



INTERNET  
SCIENTIFIC  
PUBLICATIONS

The Internet Journal of Dental Science 2009 : Volume 7 Number 1

# **Obstructive Sleep Apnoea: Dental Implications & Treatment Strategies**

**Deshpande Sae S M.D.S.**

Sr. Lecturer

Department of Prosthodontics

VSPM's Dental College & Hospital

**Soumil Sarin P M.D.S.**

Senior Resident

Department of Prosthodontics

Post-Graduate Institute, Chandigarh

**Sundarkar Pravin M.D.S.**

Associate Professor

Department of Prosthodontics

VSPM's Dental College & Hospital

**Citation:** D. Sae S, S. Sarin P & S. Pravin : Obstructive Sleep Apnoea: Dental Implications & Treatment Strategies. *The Internet Journal of Dental Science*. 2009 Volume 7 Number 1

**Keywords:** obstructive sleep apnoea | prosthetic appliances

## Abstract

The management and treatment of patients who suffer from snoring and obstructive sleep apnoea (OSA) have long since been addressed by medical and specialist dental publications and were recently highlighted by a number of mainstream dental articles. These articles inform of the medical implications of the problem and the medical and dental treatment options available, but provide little technical information on constructing these appliances. The article reviews the concepts, rationale and technical aspects of appliance design and discusses clinical and financial issues.

---

## Rationale And Concepts Of Appliance Design

Snoring is a sign of partial upper airway obstruction during sleep. Snoring and OSA are caused by abnormal airway (base of tongue and soft palate) anatomy and altered respiratory control mechanisms.

Dental appliances may prevent snoring and OSA by modifying the position of the upper airway structures so as to enlarge and/or reduce collapsibility of the airway.<sup>1,2,3,4,5</sup>

Both the superior airway space (between the soft palate and posterior nasopharynx)

and the posterior airway space (between the base of the tongue and posterior oropharynx) may be increased.<sup>6</sup> Three dimensional reconstructions of computed tomography and magnetic resonance imaging scans demonstrate significant increases in airway dimensions with the appliances.<sup>7,8</sup> The effectiveness of appliance therapy, however, depends on the severity of the sleeping disorder, the airway anatomy and whether the patient can tolerate the appliance. It is generally

advocated for mild OSA and simple snoring (i.e. snoring in the absence of OSA) and moderate to severe OSA as an alternative to nasal continuous positive airway pressure (nCPAP) or craniofacial surgery.<sup>9</sup> The use of intra-oral appliances is simple, non-invasive, reversible and cost effective and may be the basis of definitive lifelong treatment.

Originally there were three concepts for a dental appliance to modify the airway, which could be used alone or in combination depending on where the airway obstruction occurred:

**Soft palate lifting** – the prosthesis lifts and/or stabilizes the soft palate, preventing vibration during sleep.

**Tongue retention** – tongue-retaining devices (TRDs) incorporate an anterior hollow bulb, which generates a negative pressure vacuum when the tongue is inserted. The tongue is held forward, away from the posterior pharyngeal wall, opening up the airway. Owing to muscle anatomy, this appliance simultaneously modifies the position of the mandible.

**Mandibular repositioning** – these appliances (MRAs) hold the mandible in an anteroinferior position, which, as a consequence of muscle attachment, indirectly brings the tongue forward, opening up the posterior airway (Figure 1). The repositioning may also stretch and reduce the collapsibility of the soft palate via its connection to the base of the tongue and increase the superior airway space.

The soft palate lifting design is not often used because of patient tolerance and the fact that tongue posture rather than soft palate position is considered to have a significant influence on the patency of the upper airway.<sup>10</sup> An example of this type of appliance is clearly shown in the paper by Clark and Nako.<sup>5</sup> Whilst TRDs directly move the tongue to open the airway, patient tolerance is not satisfactory. When appliances are indicated the consensus appears to be a type of MRA, which Johal and Battagel consider most suitable for obstructions at the level of the tongue.<sup>2</sup> However, there is considerable variation with respect to the vertical and horizontal planes of mandibular repositioning and the material used for construction.

## **Mandibular Repositioning Considerations**

Appliances for the treatment of snoring and OSA have been described with different degrees of horizontal and vertical repositioning. The most common mandibular repositioning dimension quoted is 50–75% of maximal protrusion (approximately 5–7 mm) with minimal vertical opening.<sup>11,12,1,5</sup> The rationale for minimal opening is that, as the mandible opens, it rotates in an inferior and posterior direction. Concurrent posterior movement of the tongue and soft palate with wider opening may narrow the pharyngeal airway.<sup>7</sup> An extensive review led by Ivanhoe found a variety of protrusive dimensions were associated with successful outcomes.<sup>13</sup> This, along with Lowe's text<sup>14</sup> suggests that most protrusive positions are effective.

## **Material Considerations**

To date either poly-vinyl vacuum-formed thermoplastic materials (soft or stiff blanks depending on the Shore-hardness) or hard acrylic (predominantly heat cured) can be used. The choice of material often depends on appliance design and operator preference. To optimize retention for the vinyl appliances it is recommended that the flanges extend at least 3 mm past the gingival margin but, if softer vinyl is used, this depth should be increased.

## **Occlusal coverage**

The designs in Table 1 have varying occlusal coverage but ideally complete occlusal coverage should be prescribed to prevent the potential of localized tooth over-eruption. Case reports of complications occurring when full coverage was not used are discussed below.

## **Edentulous Patients**

Most appliances are made for dentate mouths but a technique for edentulous sufferers has been described.<sup>15</sup> If retention is a problem then a TRD may be an alternative option.

## **Some Designs Of MRAS (Table 1)**

Appliances can be considered as either one- or two-piece designs. One-piece designs are not adjustable whilst the two piece appliances can be adjusted in the anteroposterior plane. They may also permit, as with the Herbst-type appliance, limited lateral movements for parafunctional patients.

Appliance name	Description	Type	Occlusal coverage
<b>Adjustable Soft Palate Lifter</b>	Lifts the soft palate and prevents vibration during sleep	One-piece	Full
<b>Equaliser</b>	Repositions the mandible anteriorly, equalizes intra- and extra-oral air pressure (via tubes) and elevates soft palate	One-piece	Full
<b>Herbst</b>	Repositions the mandible anteriorly, in open and closed position, with adjustable struts.	Two-piece	Full
<b>Jasper Jumper</b>	Repositions the mandible anteriorly	Two-piece	Full
<b>Mandibular Repositioner</b>	Repositions the mandible anteriorly (3–7 mm)	One-piece	Full
<b>Nocturnal Airway Patency Appliance</b>	Repositions the mandible anteriorly and inferiorly	One-piece	Full
<b>Sleep and Nocturnal Obstructive Apnoea Reducer (SNOAR)</b>	Repositions the mandible anteriorly (6–9 mm) and inferiorly (17 mm)	One-piece	Excludes anterior maxilla
<b>Snore Guard</b>	Repositions the mandible anteriorly (3 mm behind maximal protrusion) and inferiorly (7 mm)	One-piece	Excludes molars
<b>Snoring Treatment Appliance</b>	Repositions the mandible anteriorly	One-piece	Excludes molars
<b>Tepper Oral Proprioceptive Stimulator</b>	Modifies tongue position	One-piece	Not applicable
<b>Tongue Locking Device</b>	Prefomed elastic appliance with a cavity that holds the tongue forward by self-created vacuum	One-piece	Full
<b>Tongue Positioner and Exerciser</b>	Helps train the tongue to be in a more favourable position	One-piece	Full
<b>Tongue Retaining Device</b>	Holds the tongue anteriorly during sleep by negative pressure in a bulb	One-piece	Full
<b>Therasnore</b>	Lower fin prevents tongue and mandible dropping back	Two-piece	Full

**Table 1. A summary of dental appliances described for the treatment of snoring and obstructive sleep apnoea**

One of the most accepted designs is the one-piece non-adjustable soft vinyl vacuum-formed MRA. (Fig 1) This involves fusing the thermoplastic material covering the maxillary and mandibular arches in the desired anteroinferior position. The occlusal position is established and recorded *in vivo* by a variety of methods including a wax-bite, silicone-bite or anterior jig with an interocclusal registration.<sup>1,4</sup> An optional anterior breathing hole can be placed to reassure those patients who mouth breathe but is not essential.




Figure 1. Conventional vinyl MRA (nonadjustable).

In the two-part Herbst-based IST (Intra- Oral Snoring Therapy) and Silensor appliances (Fig 2), the arches are connected by pivoting bars that can be altered in length to titrate the protrusive mandibular position for effectiveness and comfort. The occlusal registration for these two-part appliances is not as important because mandibular repositioning can be titrated from the intercuspal position.



**Figure 2. The Silensor – an adjustable two-part**

Although designed for orthodontic treatment, the Twin-Block functional appliance can be used as a MRA because its (two-piece adjustable) design allows adjustment of the mandibular position by removing or adding acrylic. These adjustments will be difficult, however, if modifications have been included in an attempt to prevent nocturnal mandibular opening .



**Figure 2. IST Appliance – an adjustable two-part Herbst-type appliance.**

It is crucial that full coverage is provided to prevent over-eruption and the technician knows that the appliance is for sleep breathing disorders rather than orthodontics.

The recently advertised Therasnore Zx has a different approach, in that the appliance engages only the maxillary dentition and relies on a lingual fin to induce protrusive mandibular repositioning and/or prevent the tongue and mandible from dropping backward during sleep.<sup>16</sup> It is two-part in design and adjustable via five slots of 1.5 mm increments .



Figure 5. In vivo fitting of the Therasnore Zx appliance

It is believed that if a patient opens their mouth at night the appliance will not function. To counteract this locking, elastics are advocated for the two-part Herbst-type appliances and adequate retention of the dental arches is needed for the one-piece prostheses. However, one must question this theory as a rule. For example, the Therasnore Zx engages only the maxillary dentition and relies on the mandible being closed for the lingual fin to effect mandibular repositioning. The anterior repositioning splint (ARPS) described for temporomandibular disorders is similar to the mandibular repositioners in Table 1.<sup>18</sup> The ARPS either fits the maxillary dentition and relies on the mandible being closed or vice versa .



Figure 7. A one-piece heat-cured acrylic nonadjustable MRA (an ARPS).

## Patient Perceptions, Complications And Follow Up

Patients may recall, when questioned, the thermoplastic designs, the appliance may loosen. The ideal design is probably the more rigid acrylic two-piece adjustable Herbst-type appliance but this design was not financially viable for the

department.

Materials were sought to construct a one-piece MRA that would resist splitting.

## Summary

Aspects relating to the provision of appliances for snoring and OSA have been discussed

The clinician should decide on the most suitable design for the patient depending on factors such as diagnosis of sleep disorder, cost, parafunctional activity and occlusal registration. A formal medical assessment of the sleeping disorder is advocated before treating such patients.

## Correspondence to

Dr. Saeed Deshpande,

I-7 Sagar Palace,

Laxminagar,

Nagpur – 22,

India.

Ph. No.: (Mob.) +91 9225213204

Ph. No.: (Res.) +91 712 2221830

E-mail Id: drsaeed@sify.com

## References

1. Lyons MF, Cameron DA, Banham SW. Snoring, sleep apnoea and the role of dental appliances. *DentUpdate* 2001; 28: 254–256. (s)
2. Johal A, Battagel JM. Current principles in the management of obstructive sleep apnoea with mandibular advancement appliances. *Br Dent J* 2001; 190: 532–536. (s)
3. Friedlander AH, Friedlander IK, Pagrel MA. Dentistry's role in the diagnosis and co-management of patients with sleep apnoea/hypopnoea syndrome. *Br Dent J* 2000; 189: 76–80. (s)

4. Cameron DA, Lyons MF, Fox DL, Banham SW. Pilot study of a semi-flexible intra-oral appliance for the control of snoring. *Br Dent J* 1998; 185: 304–307. (s)
5. Johal A. The relationship between obstructive sleep apnoea and dentistry: 2. Management. *Dent Update* 1998; 25: 474–477. (s)
6. Schmidt-Nowara W, Lowe A, Wiegand L, Cartwright R, Perez-Guerra F, Menn S. Oral appliances for the treatment of snoring and obstructive sleep apnoea: a review. [An American Sleep Disorders Association Review.] *Sleep* 1995; 18: 501–510. (s)
7. Lowe AA, Gionhaku N, Takeuchi K, Fleetham JA. Three-dimensional CT reconstructions of tongue and airway in adult subjects with obstructive sleep apnoea. *Am J Orthod Dentofac Orthop* 1986; 90: 364–374. (s)
8. Ryan CF, Lowe AA, Li D, Fleetham JA. Magnetic resonance imaging of the upper airway in obstructive sleep apnoea before and after chronic nasal continuous positive airway pressure therapy. *Am J Respir Crit Care Med* 1991; 144: 939–944. (s)
9. American Sleep Disorders Association Standards of Practice Committee. Practice parameters for the treatment of snoring and obstructive sleep apnoea. *Sleep* 1995; 18: 511–513. (s)
10. Pae EK, Lowe AA, Sasaki L, Price C, Tsuchiya M, Fleetham JA. A cephalometric and electromyographic study of upper airway structures in the upright and supine position. *Am J Orthod Dentofac Orthop* 1994; 106: 52–59. (s)
11. Lowe AA. Dental appliances for the treatment of snoring and obstructive sleep apnoea. In: Kryger M, Roth T, Dement W, eds. *Principles and Practice of SleepMedicine*, 2nd ed. Philadelphia: WB Saunders Co., 1994; pp.722–735. (s)
12. Clark GT, Nakono M. Dental appliances for the treatment of obstructive sleep apnoea. *J Am Dent Assoc* 1989; 118: 611–617. (s)
13. Ivanhoe JR, Cibirka RM, Lefebvre CA, Parr GR. Dental considerations in upper airway sleep disorders: A review of the literature. *J Prosthet Dent* 1999; 82: 685–698. (s)

14. Lowe AA. Dental appliances for the treatment of snoring and obstructive sleep apnoea. In: Kryger M, Roth T, Dement W, eds. *Principles and Practice of SleepMedicine*, 3rd ed. Philadelphia: WB Saunders Co., 2000; pp.929–939. (s)
  15. Meyer JB, Knudson RC. Fabrication of a prosthesis to prevent sleep apnea in edentulous patients. *J Prosthet Dent* 1990; 63: 448–451. (s)
  16. Therasnore Zx official documentation pack, 2001. 'Stop snoring and wake up feeling more rested and energetic'. 'Fitting the adjustable Zx antisnoring appliance – a guide for dentists and other professions'. [www.zxsnoreless.com](http://www.zxsnoreless.com). (s)
  17. Liu Y, Zeng X, Fu M, Huang X, Lowe AA. Effects of mandibular repositioner on obstructive sleep apnoea. *Am J Orthod Dentofac Orthop* 2000; 118: 248–256. (s)
  18. Gray RJM, Davies SJ, Quayle AA. *Temporomandibular Disorders: A Clinical Approach*. London: British Dental Journal, 1995; p.43. (s)
- 

This article was last modified on Sun, 19 Jul 09 16:55:53 -0500

This page was generated on Mon, 01 Mar 10 12:40:00 -0600, and may be cached.